

STATEMENT OF COMPLIANCE

We, OMRON Healthcare Co., Ltd, as manufacturer, hereby declare, that the products covered in Appendix A of this document are in conformance with the applicable security requirements in Schedule 1 and 2 of The Product Security and Telecommunications Infrastructure (Security Requirements for Relevant Connectable Products) Regulations 2023 (“Security Requirements”).

Password

Password policy for OMRON Connect:

1. Password Length: The user’s password must be between 8 and 20 characters long.
2. Character Requirements: Your password must contain at least one of each of the following:
 - One number (0-9)
 - One uppercase letter (A-Z)
 - One lowercase letter (a-z)
 - One special character (such as!"#\$%&'()+,-./:;<=>?@[\\]^_`{|~)

The user is recommended to choose a password that is unique, not easily guessable, and not based on personal information. Avoid using common words or phrases.

Vulnerability disclosure

At OMRON, we are committed to ensuring the security and privacy of our users' data. We recognize the importance of collaboration with security researchers and members of the community to identify and address potential vulnerabilities.

If you believe you have discovered a security vulnerability within OMRON Connect or any of our associated services, we encourage you to report it to us promptly via [Contact | OMRON Global](#). Your cooperation helps us maintain the integrity of our platform and protect our users.

Updates

At OMRON, we prioritize the security and privacy of our users' data. To ensure that our platform remains secure and resilient against emerging threats, we are committed to providing regular security updates for Omron Connect(5 Years (Until 31/12/2029)). These updates will include patches, enhancements, and measures to address any identified vulnerabilities or security concerns.

Signature

This statement of compliance is prepared on behalf of the manufacturer of the products:

OMRON Healthcare Co., Ltd.

53 Kunotsubo , Terado-cho, Muko- shi, Kyoto Zip code:617-0002

Name: Kazunobu Itonaga

Title: Executive Officer

Place and Date: Kyoto Japan, 13 Dec 2024

Please note that this statement of compliance, including the Defined Support Period stated herein, is only applicable to products sold in the UK as described in Appendix A.



Appendix A:

Product type:	Model:	Manufacturers Name:
Upper Arm Blood Pressure monitors	HEM-7600T-E	OMRON Healthcare Co., Ltd
	HEM-7361T-EBK/ESL and	OMRON Healthcare Co., Ltd
	HEM-7381T1-EBK/EOSL	OMRON Healthcare Co., Ltd
	HEM-7155T-EBK/ESL	OMRON Healthcare Co., Ltd
	HEM-7143T1-EBK /HEM - 7143T2-ESL	OMRON Healthcare Co., Ltd
	HEM-7146T2-EBK/ESL	OMRON Healthcare Co., Ltd
	HEM-7280T-E	OMRON Healthcare Co., Ltd
	HEM-7361T-D	OMRON Healthcare Co., Ltd
	HEM-7155T-D	OMRON Healthcare Co., Ltd
	HEM-7143T1-D	OMRON Healthcare Co., Ltd
	HEM-7361T-BS	OMRON Healthcare Co., Ltd
Wrist Blood Pressure monitors	HEM-9601T-E3	OMRON Healthcare Co., Ltd
	HEM6232T-E/D	OMRON Healthcare Co., Ltd
	HEM-6161T-E/D	OMRON Healthcare Co., Ltd
OMRON Complete	HEM-7530T-E3	OMRON Healthcare Co., Ltd
Scales	HBF-222T-EBK	OMRON Healthcare Co., Ltd
	HN-300T2-EBK/EGY	OMRON Healthcare Co., Ltd
HeartGuide	HEM-6411-MAE	OMRON Healthcare Co., Ltd
Smart Thermometer Eco Temp Intelli IT	MC-280B	OMRON Healthcare Co., Ltd
Oxymeter P300 Intelli I	HPO-300T	OMRON Healthcare Co., Ltd